

Characteristics	Available Rapid HIV Tests		
	OraQuick Advance	Reveal G3 Rapid HIV-1	Multispot HIV 1/2
Manufacturer	OraSure Technologies	MedMira Laboratories	Bio-Rad
Detection	HIV-1 and -2	HIV-1	HIV-1 and -2
Specimen	Oral fluid, whole blood, plasma	Serum, plasma	Serum, plasma
Sensitivity (95%)	99.3% (oral fluid); 99.6% (whole blood <sup>b</sup> ); 99.6% (plasma)	99.8%	100%
Specificity (95%)	99.8% (oral fluid); 100% (whole blood <sup>b</sup> ); 99.9% (plasma)	99.1% (serum); 98.6% (plasma)	99.93%
Complexity <sup>a</sup>	CLIA waived for oral fluid and both venipuncture and finger stick whole blood only	Moderate Complexity: No waiver	Moderate Complexity: No waiver

CLIA, Clinical Laboratory Improvement Amendments.

<sup>a</sup> Information regarding CLIA waivers of HIV tests is available at [www.cdc.gov/hiv/topics/testing/resources/factsheets/roltCLIA.htm](http://www.cdc.gov/hiv/topics/testing/resources/factsheets/roltCLIA.htm). Information about or assistance with completing the CLIA waiver application can be obtained by calling Centers for Medicare & Medicaid Services toll-free at: 877-267-2323.

<sup>b</sup> Finger stick or venipuncture.

This publication was supported by grant number H98 HA00015 from the U.S. Health Resources and Services Administration (HRSA). This grant is funded through Part A of the Ryan White Treatment Modernization Act of 2006, through the New York City Department of Health and Mental Hygiene to Public Health Solutions of New York City, Inc. Its contents are solely the responsibility of Health Research, Inc. and do not necessarily represent the official views of the funders.

Characteristics	Available Rapid HIV Tests			
	UniGold Recombigen	Clearview HIV 1/2 Stat Pak	VITROS Anti-HIV 1+2 assay	Clearview Complete HIV 1/2
Manufacturer	Trinity Biotech	Inverness Medical	Ortho-Clinical Diagnostics	Inverness Medical
Detection	HIV-1	HIV-1 and -2	HIV-1 and -2	HIV-1 and -2
Specimen	Whole blood, serum, plasma	Whole blood, serum, plasma	Serum, plasma	Whole blood, serum, plasma
Sensitivity (95%)	100% (whole blood <sup>b</sup> ); 100% (serum and plasma)	99.7% (whole blood <sup>b</sup> ); 99.7% (serum and plasma)	100%	99.7% (whole blood <sup>b</sup> ); 99.7% (serum and plasma)
Specificity (95%)	99.7% (whole blood <sup>b</sup> ); 99.8% (serum and plasma)	99.9% (whole blood <sup>b</sup> ); 99.9% (serum and plasma)	99.58%	99.9% (whole blood <sup>b</sup> ); 99.9% (serum and plasma)
Complexity <sup>a</sup>	CLIA waived for both venipuncture and finger stick whole blood only	CLIA waived for both venipuncture and finger stick whole blood only	Moderate Complexity: No waiver	CLIA waived for both venipuncture and finger stick whole blood only

CLIA, Clinical Laboratory Improvement Amendments.

<sup>a</sup> Information regarding CLIA waivers of HIV tests is available at [www.cdc.gov/hiv/topics/testing/resources/factsheets/roItCLIA.htm](http://www.cdc.gov/hiv/topics/testing/resources/factsheets/roItCLIA.htm)  
Information about or assistance with completing the CLIA waiver application can be obtained by calling Centers for Medicare & Medicaid Services toll-free at: 877-267-2323.

<sup>b</sup> Finger stick or venipuncture.